Phone 91-7232-242456 Fax 91-7232-244148

#### SHRI. VASANTRAO NAIK GOVT. MEDICAL COLLEGE, YAVATMAL. श्री. वसंतराव नाईक शासकीय वैद्यकीय महाविद्यालय, यवतमाळ

Ref. No.SVNGMC&H/Stores/QT/21-22 /

20864

# Open Quotation Notice (प्रथम महनेवाह

## For the supply of instrument for Dermatology, Venereology and Leprosy.

Dean Shri Vasantrao Naik Govt. Medical College, Yavatmal invited separate open quotation for the supply of Instruments listed below in two envelope system from manufacturer & authorized dealers. Detail list & specification is enclosed. Interested Manufacturer/dealer can send their quotation.

- ENVELOPE NO:-1 TECHNICAL BID:- Following documents must be submit in this envelope
- Shop/Udhyog Aadhar/GEM Reg. establishment act registration. 1.
- PAN Card
- GST Registration certificate
- Cancel Cheque & Bank Details on letter head.
- Acceptance of all terms of conditions with stamp of signature firm/company on letter head.
- Authorization certificate from the manufacturer 6

#### ENVELOPE NO :- 2 PRICE BID B.

- 1. Quote your Minimum Rates in a following Format on the letterhead of the firm/company in Price Bid envelop the cost should be quoted in Indian currency only. Quote maximum two models on separate sheet.
  - 1.Basic Price:-....
  - 2. Taxes: Excise duty..... Custom duty..... GST..... other......
  - 3. Installation Charges:-....
  - 4. TOTAL=(i)+(ii)+(iii)+(iv):-....

  - 5. CMC/AMC:-.... 6. Prices for consumables/ Disposable/ accessories should be submitted separately in price bid which will be valid for warranty period. AMC.CMC period.

Both the sealed envelope No. 1 and No.2 should be put together in one cover and it should be sealed. This sealed envelope should be marked on upper left hand corner mentioning name of the items No of the Item Quotation enquiry No. And Date of quotation opening. The full name and address of the Firm/Company should be written on the left hand lower corner. Such sepa rate envelope containing sets of Envelope No.1 and No. 2 should be submitted separately for each Items mentioned in the List.

Envelope No.2 will be opened only when the documents as mentioned in envelope No.1 are in order as per requirements and if item is as per specifications.

Separate quotation should be prepared for every item mentioned in the list and sent in sealed envelope (duly sealed) by post or hand delivery so as to reach Inward Section of administrative office. Shri Vasantrao Naik Govt. Medical College, Yavatmal on or before 17-1/2022 up to 05.00 pm.

Date & timing of the opening of quotation (Technical Bid):- 18 / 01 /2022 at 12.00 pm. at College Council Hall. Date & time of opening of the Price bid will be decide after scrutiny of the technical detail. Representative of the Supplier may attend the meeting by submitting authorization letter.

Please quote your minimum rates for Supply of articles as per details given below/list enclosed. The separate envelopes (every Item) containing the quotation should be sealed and super-scribed as under:-

Quotation due on...... Against inquiry No...... for the purchase of......

LIST OF INSTRUMENT REQUIRED

Sr	Name of the Instrument/Equipment & Specification	QTY
No	Name of the instrument Equipment & Specification	4.
1	Cryo-Gun Liquid Nitrogen .	. 01
2	Cryo- can 3.9 Litres	01
3	Portable hand Foot PUVA	01
4	Magnifying Lens with LED Lamp (Handheld with self illumination)	0.5
5	Bifrecator Cautery Machine	01
6	lontophoresis machine	01,
7	Suction Blister Machine	.01
8	Surgical Smoke Evacuator	. 01
9	Multimedia Projector	01
10	Disposable skin Biopsy punches 2mm,	20
	2.5mm,	. 20
	3mm,	100
	4mm.	
11	Nail Cutter:	
	Staight	01
	curved	
12	Nail Extractor Dissector:	03.
	3 Different Sizes	
13.	Nail Spatula Chisel Ended	0.1
14	hand Dermabrasser 10mm	01
15	skin Grafting Humby's Knife (3-5 tinches)	01
16	Comedo extractor-	
	Double fine special Oval cup:	
	3-4mm	'02'
	4-5mm	02
	5-6mm	02
17	Molluscum Curette:	
	1-2 mm	02
	2-3mm	02
	3-4mm	.02

The terms and conditions for the supply are:-

1. Enquiry No. must invariably be given at the top of the envelope. Envelopes without indication will not be entertained.

2. For equipment/ accessories/ spares, technical bid and price bid must be sent in two separate, sealed envelopes and front of the envelop marked clearly in capitals as 'PRICE BID', 'TECHNICAL BID'. Otherwise the bid will be summarily REJECTED.

3. Offers other than those from the manufacturer should be supported by an authority letter from the manufacturer as per Annexure-A.

4. AMC/CMC/Extended Warranty if applicable must be quoted along with price bid and not with the Technical bid. However, in technical bid form consent for AMC/CMC must be given.CMC (Labor & Spare parts), these charges include the cost of replacement parts/Components/spares/accessories etc. This shall be borne by the firm. In case of consumables/these will also be included in CMC as per Technical Specification. However, AMC charge will cover labor only and does include the cost of any replacement parts/components etc.

#### 5. Unsealed quotations will be rejected.

6. Full description, make and quantity of the articles must be mentioned in your quotation. Original catalogue/literature should be submitted with the quotation (Technical bid envelop)

- 8. Delivery period for each item be specifically mentioned in the quotations from the date of Supply order it should be maximum of 6 weeks.
- 9. The Dean reserves the right to call the firm for physical demonstration of item to verify the quality of the offered item from the lowest bidder. The firm must demonstrate the product within 7 days of requisition. failing which the bid will be rejected.
- 10. The institute reserves the right to reject the supplied goods if the same are not found in accordance with the specifications approved and asked for. In case there is a short/defective supply the firm will be informed and the defective material will be lifted from the concerned department/Central store by the supplier at their own cost within two weeks period. The institute will not bear any expenses on this account.
- 11. The rates and total cost must be quoted in both words and figures (over-writing not allowed.)
- 12. Rate should be for free delivery at Shri V.N. Govt. Medical College, Yavatmal premises only. Rate should be quoted including all charges. Rates must be quote on letter head alongwith Pan. VAT. CST/GST Number. It should be submit in Price envelope. (envelope No 2)
- 13. The suppliers will have to adhere to the delivery schedule. Delayed supplies beyond 30 days from the date of supply order as the case may be will be subject to demurrage @0.5% per weeks of part thereof of the contract price subject to maximum of 10% of contract price, if the Price is below 2 lakh, else it will be @0.5% up to as maximum of 5%.
- 14. Warranty/Guarantee should be quoted as per the Technical specifications. (2 Years)
- 15. The successful bidder will submit P.B.G in the shape of FDR/TDR only in favor of **Dean**, **Shri Vasantrao Naik Government Medical College**, **Yavatmal** @ 3% of the total value of the order which will be valid for 30 months in case the warranty is for 24 months and 66 months in case the warranty is for five years. The same shall be further extended for 5 more year till the time AMC/CMC inforce.
- 16. The Manufacturer/Distributor/Agent should furnish along with the Price Bid a certificate that the rates quoted by the firm are the same as are being charged from other Government or Non-Government Departments/Institutions for similar items.
- 17. Compliance Chart: (If quoting for more than one model then separate formats should be enclosed for different models in the same envelope.) Make & Model No of the quoted Item should be mentioned on the top Part of the compliance Chart.

Sr No	Tender/Quotation specifications as asked in the tender form	Equivalent specifications quoted by the tendered with the name of the manufacturer	Whether the tender/ Quotation quoted by the supplier is as per specifications asked for Indicate against each item clearly YES/NO	If the tender quoted is not as per specifications then variations/deviations should be clearly indicated against each item asked for.
1	2	3	4	5

,
of have any direct

or indirect interests with purchasing institute.

19. The Envelop No. 2 (Commercial Bid) will be opened of only those tenderers who have submitted all the documents as per tender terms and conditions in the envelop no. 01 and are technically qualified. The demonstration of tenderer who had quoted lowest price will be taken. If two tenderers had quoted same cost of equipment, then the demonstration of both the tenderer will be taken.

20. Complaint/s:—Any complaint / representation regarding tender/quotation will be entertained only after depositing of Rs. 10,000/- in form of Demand Draft in the name of Dean Shri V N Govt Medical College Yavatmal. Subsequently necessary action will be taken by the Dean and decision of Dean will be binding upon the complainant. If the complaint turns out to the false or invalid the amount will be forfeited. The Amount shall be refunded if after scrutiny the complaint is found to be true. No further complaint/ representation from the same complainant for the same tender/quotation will be entertained.

Note:- The Technical Specification, quoted price & availability of service Centre near Yavatmal will be considered while approving the prices.

21. 1) The Instrument/Equipment Shall be of ISI/ISO mark or equivalent.

II) It shall necessary to provide OEM (original Equipment Manufacturer)

III) Supplier to provide user of maintenance AMC/CMC additional services.

Shri Vasantrao Naik Government Medical

College, Yavatmal

Harry 12 Car

#### **Technical Specifications of Cryo-gun Liquid Nitrogen**

- 1. It should have four spray tips with apertures of various sizes (0.04-0.016 inches), a 20 gauge straight and 20 gauge bent spray for more accurate and controlled treatment.
- 2. Its fingertip trigger should provide maximum control and easy viewing, it should rotate for left hand users.
- 3. It should have safety auto-vent for gradually venting internal pressure when cap is unscrewed.
- 4. Its relief valve should maintain constant operating pressure for consistent and accurate freezing.
- 5. It should have stainless steel & brass construction, polypropylene base, polypropylene cover & collar.
- 6. It should have Capacity: 10 oz (300ml) and Static holding time: 10 to 12 hours.

#### **Technical Specifications of Cryo-can**

- 1. It should be high quality stainless steel Devar flask with elegant & stable base.
- 2. It should be light weight and easy to use.
- 3. It should have capacity of 3.7 litres.

प्राध्यापक व विभाग प्रमुख त्वचारोग व गुप्तरोग विभाग

श्री व.ना.शा.वै. महाविद्यालय, यत्तवनारु

#### Technical Specifications of Hand-foot UVA & UVB single panel

- 1.It should have sincrolite sheet ensuring uninterrupted irradiation.
- 2. It should have Phillips tubes TL/10R for UVA & TL/01 for NB UVB, each 2 tubes.
- 3. It should mirror type reflector ensuring maximum irradiation.
- 4. Specially designed patient data cards should be provided with it.

प्राध्यापक व विभाग प्रमुख त्वचारोग व गुप्तरोग विभाग श्री व.ना.शा.वै. महाविद्यालय, यतवमाळ

## Technical Specifications of Magnifying Lens with LED lamp (handheld with self-illumination)

- 1. It should work on electrical supply of 12Vand not be battery operated.
- 2. It should be portable, handheld device.
- 3. It should have self-illumination by LED lamp for clear view.
- 4. It should have magnification power of 6x.
- 5. It should be electric shock-proof, compact & of light weight.
- 6. It should have built in choke (shock proof).

प्राध्यापक व विभाग प्रमुख त्वचारोग व गुप्तरोग विभाग श्री व.ना.शा.वै. महाविद्यालय, यतवनाक

## **Technical Specifications of Byfrecator Cautery machine**

- 1. It should be portable.
- 2. It should have precisely variable intensity upto 50 Watts.
- 3. It should have bipolar cable.
- 4. It should offer instant electro-coagulation, electro-dessication, electro-fulgration & electrosection.

प्राध्यापक व विभाग प्रमुख त्वचारोग व गुतरोग विभाग श्री व.ना.शा.वै. महाविद्यालय, यतवसाक

#### Technical Specifications Of IontophoresisMachine

- 1. It should have adjustable current strength.
- 2. It should have main plugof 230 V/110V input.
- 3. It should be electric shock-proof, compact &of light weight.
- 4. It should have carrying case.
- 5. It should have big non-corrosive metal electrode plates, a pair thick connection cables with 4mm standard banana plug and two plastic water trays.

6. It should have bright digital display for easy monitoring & controlling output current during treatment.

प्राध्यापक व विभाग प्रमुख त्यचारोग व गुप्तरोग विभाग श्री व.ना.शा.वै. महाविद्यालय, यतवनाक

#### **Technical Specifications of Suction Blister Apparatus**

- 1. It should be portable.
- 2. it should be able to create prolonged suction at a negative pressure of -200 to -500mmHg to the donor site

3. it should be provided with suction cups of different sizes.

प्राध्यापक व विभाग प्रमुखें त्वचारोग व गुप्तरोग विभाग

श्री व.ना.शा.वै. महाविद्यालय, यततमाल

#### **Technical Specifications of Multimedia projector**

- 1. It should create big screen with bright ambient light, equally high white and colour light output.
- 2. It should be portable and easy to set up.
- 3. Its Projection system should have 3LCDtechnology, RGB liquid crystal shutter, LCD panel of 0.55 inch with D7.
- 4. It should have Colour Light Output of 3,300 Lumen- 2,050 Lumen In accordance with IDMS15.4 and White Light Output of 3,300 Lumen 2,050 Lumen In accordance with ISO 21118:2012.

5.It should have Resolution: SVGA, 800 x 600, 4:3,

Aspect Ratio: 4:3, Contrast Ratio: 15,000: 1

Light source: LampUHE, 210 W, 6,000 h durability, 10,000 h durability

Keystone Correction: Auto vertical: ± 30 °, Manual horizontal ± 30 °

Colour Processing: 10 Bits, 2D Vertical Refresh Rate: 100 Hz - 120 Hz

Colour Reproduction: Upto 1.07 billion colours

6. It should have Optical Projection Ratio: 1.45 - 1.96:1, Zoom: Digital, Factor 1.35.

Lens: Optical, Image Size: 30 inches - 350 inches

Projection Distance Wide/Tele: 1.77 m - 2.4 m ( 60 inch screen)

Projection Lens F Number should be 1.44

Focal Distance should be 16.7 mm, Focus Manual Offset should be 8: 1

7. It should have ConnectivityUSB Display Function: 2 in 1: Image / Mouse, Interfaces should be USB 2.0 Type A, USB 2.0 Type B, VGA in, HDMI in, Composite in, Cinch audio in, Wireless LAN IEEE 802.11b/g/n (optional)

8. It should have Energy Use: 282 Watt, 203 Watt (economy), 0.2 Watt (standby), On mode power consumption as defined in JBMS-84 257 Watt, Supply Voltage: AC 100 V - 240 V, 50 Hz - 60 Hz.

प्राध्यापक व विभाग प्रमुख त्वचारोग व गुप्तरोग विभाग श्री व.ना.शा.वै. महाविधालय, यतवमाक

## **Technical Specifications of Surgical Smoke evacuator**

1. It should be portable.

2. It should have pre filter, carbon filter & HEPA filter.

3. It should have Airflow: 20m³/min

Suction pressure: 100mm Hg

Suction power: 800W

Filtration stage: 3 stage filter (pre+carbon+HEPA H13)

Electrical: 100 -240 V, 50/60 Hz.

प्राध्यापक व विभाग प्रमुख त्वचारोग व गुप्तरोग विभाग श्री व.ना.शा.वै. महाविद्यालय, यतवमाळ