

E- Open Quotation Notice

For the supply of Equipment & CAL Software

Dean Shri Vasant Rao Naik Govt. Medical College, Yavatmal invited separate open quotation for the supply of Instruments **listed below** in **two envelope** system from manufacturer & authorized dealers. Detail list & specification is enclosed. Interested Manufacturer/dealer can send their quotation.

A. ENVELOPE NO :- 1 TECHNICAL BID:- Following documents must be submit in this envelope

1. Shop/Udhyog Aadhar/GEM Reg. establishment act registration.
2. PAN Card
3. GST Registration certificate
4. Cancel Cheque & Bank Details on letter head.
5. Acceptance of all terms of conditions with stamp of signature firm/company on letter head.

B. ENVELOPE NO :- 2 PRICE BID

1. Quote your Minimum Rates in a following Format on the letterhead of the firm/company in Price Bid envelop the cost should be quoted in Indian currency only. Quote maximum two models on separate sheet.

1. Basic Price:-.....

2. Taxes: Excise duty..... Custom duty..... GST..... other.....

3. Installation Charges:-.....

4. TOTAL=(i)+(ii)+(iii)+(iv):-.....

5. CMC/AMC:-.....

6. Prices for consumables/ Disposable/ accessories should be submitted separately in price bid which will be valid for warranty period, AMC.CMC period.

Both the sealed envelope No. 1 and No.2 should be put together in one cover and it should be sealed. This sealed envelope should be marked on upper left hand corner mentioning name of the items No of the Item Quotation enquiry No. And Date of quotation opening. The full name and address of the Firm/Company should be written on the left hand lower corner. Such separate envelope containing sets of Envelope No.1 and No. 2 should be submitted separately for each Items mentioned in the List.

Envelope No.2 will be opened only when the documents as mentioned in envelope No.1 are in order as per requirements and if item is as per specifications.

Separate quotation should be prepared for every item mentioned in the list and sent in sealed envelope (duly sealed) by post or hand delivery so as to reach Inward Section of administrative office, Shri Vasant Rao Naik Govt. Medical College, Yavatmal on or before **06 /07 /2024** up to 05.00 pm.

Date & timing of the opening of quotation (Technical Bid) :- **08 / 07 /2024 at 12.00 pm.** at College Council Hall. Date & time of opening of the Price bid will be decide after scrutiny of the technical detail. Representative of the Supplier may attend the meeting by submitting authorization letter.

Please quote your minimum rates for Supply of articles as per details given below/list enclosed,. The separate envelopes (every Item) containing the quotation should be sealed and super-scribed as under:-

Quotation due on..... Against inquiry No.....for the purchase of.....

LIST OF INSTRUMENT REQUIRED

Sr No	Name of the Instrument/Equipment & Specification	Approx QTY	Approx Unit Price inclusive all Taxes
1	CAL Software	01	
2	Automatic Isolated Organ Bath	01	
3	Elisa Reader	01	
4	Vertical autoclave (40 Liters	01	

The terms and conditions for the supply are:-

1. Enquiry No. must invariably be given at the top of the envelope. Envelopes without indication will not be entertained.
2. For equipment/ accessories/ spares, technical bid and price bid must be sent in two separate, sealed envelopes and front of the envelop marked clearly in capitals as 'PRICE BID', 'TECHNICAL BID'. Otherwise the bid will be summarily REJECTED.
3. Offers other than those from the manufacturer should be supported by an authority letter from the manufacturer as per Annexure-A.
4. AMC/CMC/Extended Warranty if applicable must be quoted along with price bid and not with the Technical bid. However, in technical bid form consent for AMC/CMC must be given.CMC (Labor& Spare parts), these charges include the cost of replacement parts/Components/spares/accessories etc. This shall be borne by the firm. In case of consumables/these will also be included in CMC as per Technical Specification. However, AMC charge will cover labor only and does include the cost of any replacement parts/components etc.

5. Unsealed quotations will be rejected.

6. Full description, make and quantity of the articles must be mentioned in your quotation. Original catalogue/literature should be submitted with the quotation (Technical bid envelop)
7. Validity of the quotation should be for a minimum period of 120 days.
8. Delivery period for each item be specifically mentioned in the quotations from the date of Supply order it should be maximum of 6 weeks.
9. The Dean reserves the right to call the firm for physical demonstration of item to verify the quality of the offered item from the lowest bidder. The firm must demonstrate the product within 7 days of requisition, failing which the bid will be rejected.
10. The institute reserves the right to reject the supplied goods if the same are not found in accordance with the specifications approved and asked for. In case there is a short/defective supply the firm will be informed and the defective material will be lifted from the concerned department/Central store by the supplier at their own cost within two weeks period. The institute will not bear any expenses on this account.

11. The rates and total cost must be quoted in both words and figures (over-writing not allowed.)

12. Rate should be for free delivery at Shri V.N. Govt. Medical College, Yavatmal premises only. Rate should be quoted including all charges. Rates must be quote on letter head alongwith Pan, VAT, CST/GST Number. It should be submit in Price envelope. (envelope No 2)
13. The suppliers will have to adhere to the delivery schedule. Delayed supplies beyond 30 days from the date of supply order as the case may be will be subject to demurrage @0.5% per weeks or part thereof of the contract price subject to maximum of 10% of contract price, if the Price is below 2 lakh, else it will be @ 0.5% up to as maximum of 5%.
14. Warranty/Guarantee should be quoted as per the Technical specifications. (2 Years)

15. The successful bidder will submit P.B.G in the shape of FDR/TDR only in favor of **Dean, Shri Vasantrao Naik Government Medical College, Yavatmal @ 3%** of the total value of the order which will be valid for 30 months in case the warranty is for 24 months and 66 months in case the warranty is for five years. The same shall be further extended for 5 more year till the time AMC/CMC inforce.

16. The Manufacturer/Distributor/Agent should furnish along with the Price Bid a certificate that the rates quoted by the firm are the same as are being charged from other Government or Non-Government Departments/Institutions for similar items.

17. Compliance Chart :- (If quoting for more than one model then separate formats should be enclosed for different models in the same envelope.) Make & Model No of the quoted Item should be mentioned on the top Part of the compliance Chart.

Sr No	Tender /Quotation specifications as asked in the tender form	Equivalent specifications quoted by the tendered with the name of the manufacturer	Whether the tender/ Quotation quoted by the supplier is as per specifications asked for Indicate against each item clearly YES/NO	If the tender quoted is not as per specifications then variations/deviations should be clearly indicated against each item asked for.
1	2	3	4	5

Equipment Name - _____ Make & Model: _____

18. The firm should submit undertaking on letter head stating that firm & its personnel do not have any direct or indirect interests with purchasing institute.

19. The Envelop No. 2 (Commercial Bid) will be opened of only those tenderers who have submitted all the documents as per tender terms and conditions in the envelop no. 01 and are technically qualified. The demonstration of tenderer who had quoted lowest price will be taken. If two tenderers had quoted same cost of equipment, then the demonstration of both the tenderer will be taken.


20. Complaint/s :- Any complaint / representation regarding tender/quotation will be entertained only after depositing of **Rs. 10,000/-** in form of **Demand Draft** in the name of Dean Shri V N Govt Medical College Yavatmal. Subsequently necessary action will be taken by the Dean and decision of Dean will be binding upon the complainant. If the complaint turns out to be false or invalid the amount will be forfeited. The Amount shall be refunded if after scrutiny the complaint is found to be true. No further complaint/ representation from the same complainant for the same tender/quotation will be entertained.

Note:- The Technical Specification, quoted price & availability of service Centre near Yavatmal will be considered while approving the prices.

21. I) The Instrument/Equipment Shall be of ISI/ISO mark or equivalent.

II) It shall necessary to provide OEM (original Equipment Manufacturer)

III) Supplier to provide user of maintenance AMC/CMC additional services.


DEAN

**Shri Vasantrao Naik Government Medical
College, Yavatmal**



Annexure 1: Automatic isolated organ bath
Specifications

Single unit with digital temperature controller: should include the following

Organ bath tube PP(Perspex material)

Organ bath tube glass(inner organ bath tube of glass material)

Reservoir coil

Copler set

Adequate Silicon tube for interconnections

Aerator

Adequate Tubing for aerator

Ink marker

Frontal lever for holding ink marker

Marriot bottle 1 ltr

Stand for Marriott bottle

Stirrer

Digital thermostat with temperature range of 30° C to 60° C.

Professor And Head
Dept.Pharmacology
S.V.N.G.M.C. & H. Yavatmai

Annexure 2: Computer Assisted Learning(CAL) software



CAL software: specifications

Should meet the following conditions:

- 1) It should be possible to install the software in 05 computers
- 2) Software should run in offline mode
- 3) Life-time subscription (with free updates, if any)
- 4) Satisfactory pre-purchase as well as post-purchase demo. **(Order will be placed only after satisfactory pre-purchase demo)**
- 5) Should have practice mode and exam mode (videos not required, as already available)


The software should have the following experiments: interactive assessment modules

I) Pharmacodynamics - Testing of Effect of Drug (In Vitro)

1. Dose Response Curve - Using Guinea Pig ileum
2. Bioassay of Histamine using Guinea Pig Ileum

II) Pharmacodynamics - Testing of Effect of Drug (In Vivo)

3. Effect of Drugs on Blood Pressure - Dog Blood Pressure
4. Potentiation- Effect of Acetylcholine + Neostigmine on Dog BP
5. Tachyphylaxis - Effect of tyramine on Dog BP
6. Nicotinic Action of Acetylcholine on Dog BP
7. Dales Vasomotor reversal - Effect of Adrenaline and Noradrenaline on Dog BP
8. Effect of Miotics and Mydriatics on Rabbit Eye


Professor And Head
Dept. Pharmacology
S.V.N.G.M.C. & H. Yavatmal

Specification of Autoclave (40 Lit)

- The Vertical Autoclave Machine (40L) is a reliable and compact sterilization solution with a seamless stainless steel construction, this electric autoclave ensures a bacteria-free environment. The wing nut type design, along with features like a pressure gauge, safety valve, and double safety valve, makes it efficient and safe for various applications.
 - Double-wall, single-chamber design.
 - Leak-proof stainless steel construction.
 - Equipped with essential safety features.
 - Compact and portable with mild steel legs.
 - Suitable for a variety of sterilization needs.
 - Size- 12 x 22

SPECIFICATION OF ELISA READER

EIAQuant is a semi automated ELISA plate reader that uses high quality optical system and sophisticated embedded computer control technology which enables the user reliable measurement and data analysis with minimum maintenance.

- Wide Test Menu
- Touch screen display
- Built in printer
- Intuitive software
- Direct sample reports
- Assay modes – Quantitative, Semi Quantitative and Qualitative
- Microplate Types – 96 wells
- Measuring system – 8 channels fibre optics optical system
- Reading speed – Monochromatic – 6.5 sec for 96 wells
 - Bichromatic – 16 sec for 96 wells
- Light source – 12V 20W halogen lamp
- Wavelength – 405, 450, 492, 630 nm with 3 optional filter positions